




# QuantaMaster™ 60

NIR PMT Spectrofluorometer with Q-7b Second Emission Channel Upgrade

## Specifications

### NIR

Sensitivity	Singlet oxygen generated by 0.1 $\mu$ M rose bengal in MeOH
Light Source	High power continuous Xenon arc lamp
Monochromators	Czerny-Turner design
Focal Length	200 nm
Excitation Grating	1200 line/mm 300 nm blaze
Emission Grating	600 line/mm 1250 nm blaze
Optional Grating	75–2400 line/nm and holographic models available
Bandpass	0–25 nm excitation, 0–50 nm emission, continuously adjustable (computer control available)
Wavelength Accuracy	$\pm 1$ nm excitation, $\pm 2$ nm emission
Resolution	0.5 nm excitation, 1.0 emission
NIR PMT Detector	300 to 1400 nm, 950 to 1400 nm, 300 to 1700 nm or 950 to 1700 nm
System Control	Computer interface with spectroscopy software



### Second Emission Channel

Detection	Photon counting/analog
-----------	------------------------