




QuantaMaster™ 30

Phosphorescence Spectrofluorometer with Q-7p Second Emission Channel Upgrade

Specifications

Phosphorescence

Detection Limit	50 femtomolar
Minimum Lifetime	400 ns
Emission Range	185 nm to 680 nm (optional to 900 nm)
Data Acquisition Rate	0.2 points/sec to 60 points/sec
Light Source	High power Xenon flash lamp
Repetition Rate	1 to 300 Hz (software controllable)
Pulse Width	1–2 μ s
Monochromators	Czerny-Turner design
Focal Length	200 nm
Excitation Grating	1200 line/mm 300 nm blaze
Emission Grating	1200 line/mm 400 nm blaze
Optional Grating	75–2400 line/nm and holographic models available
Bandpass	0–25 nm, continuously adjustable (computer control available)
Detection	Gated (gate width and position software controlled)
Temporal Resolution	20 ns
System Control	Computer interface with spectroscopy software



Second Emission Channel

Emission Grating	1200 line/mm 400 nm blaze
Optional Grating	75–2400 line/nm and holographic models available
Bandpass	0–25 nm, continuously adjustable (computer control available)



Photon Technology International

USA:	Photon Technology International, Inc., 300 Birmingham Road, PO Box 272 Birmingham, NJ 08011 Tel: 609-894-4420, Fax: 609-894-1579, E-mail: marketing@pti-nj.com , www.pti-nj.com
Canada:	Photon Technology International, Inc., 347 Consortium Court, London, Ontario, N6E 2S8 Tel: 519-668-6920, Fax: 519-668-8437, E-mail: sales@pti-can.com
UK:	Photon Technology International, Inc., Unit M1 Rudford Industrial Estate, Ford Road, Ford, West Sussex BN180BF Tel: +44 (0) 1903 719555, Fax: +44 (0) 1903 725722, E-mail: sales@pti-uk.co.uk
Germany:	PhotoMed GmbH, Inninger Str. 1, 82229 Seefeld, Germany, Tel: +49 (0) 81 5299 3090, Fax: +49 (0) 81 5299 3098, E-mail: sales@photomed.com
Denmark:	PhotoMed GmbH, Sondre Alle 2, 4600 Koge Denmark, Tel: +45 5666 3386, Fax: +45 81 56663381, E-mail: sales@photomed.dk